

NitriDerm®

Nitrile Surgical Gloves Series 135

These nitrile synthetic surgical gloves contain no natural rubber latex proteins. Soft, highly elastic formulation provides the maximum in comfort, fit and protection, while offering exceptional tactile sensitivity.

Characteristics: Sterile, Powder-Free

SIZE	ITEM NO.
6	135260
6.5	135265
7	135270
7.5	135275
8	135280
8.5	135285
9	135290

All Sizes:

50 Gloves/Box, 4 Boxes/Case



Features:

- Flat Stacking = Gloves Dispense Easily, One by One
- Non-Latex = No Risk of Latex Allergens
- Textured Finish = Improved Wet/Dry Grip
- Low Modulus = a Softer, More Comfortable Fit

Manufactured in compliance with multiple international standards, including ASTM 3577, ASTM D5151, ASTM F1671, and ASTM D6978

Tested for use with Chemotherapy Drugs as per ASTM D6978

Standard Practice for Assessment of Medical Gloves to Permeation by Chemotherapy Drugs

Minimum Breakthrough Detection Time in Minutes:

Carmustine (3.3 mg/ml)	30.4	Fluorouracil (50.0 mg/ml)	>240
Cisplatin (1.0 mg/ml)	>240	Ifosfamide (50.0 mg/ml)	>240
Cyclophosphamide (20.0 mg/ml)	>240	Mitoxantrone (2.0 mg/ml)	>240
Dacarbazine (10.0 mg/ml)	>240	Paclitaxel (6.0 mg/ml)	>240
Doxorubicin Hydrochloride (2.0 mg/ml)	>240	Thiotepa (10.0 mg/ml)	59.0
Etoposide (20.0 mg/ml)	>240	Vincristine Sulfate (1.0 mg/ml)	>240

Caution: Testing showed an average breakthrough time of 30.4 minutes for Carmustine and an average breakthrough time of 59.0 minutes for Thiotepa.

Caution: In compliance with ASTM D6978, the testing conditions used are intended to approximate the worst-case conditions for clinical use. Testing was conducted on single layer glove material. It is the users' responsibility to determine the applicability of these gloves for their intended use with chemotherapy drugs

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	26	23
ASTM Requirement Min. (Mpa)	17	12
Elongation (%)	670	620
ASTM Requirement Min. (%)	650	490

Average Length	11.6 in ♦ 295 mm
Average Palm Thickness	5.5 mil ♦ 0.14 mm
Average Finger Thickness	6.7 mil ♦ 0.17 mm



Intertek

Innovative Healthcare Corporation is certified to ISO 13485:2016 QMS for medical devices.



**INNOVATIVE
HEALTHCARE
CORPORATION**

www.ihc-solutions.com

1-800-272-1533

Product Solutions You Trust

View this product
on our website:

