


Characteristics

Damp don gel interior provides effortless donning. Curved finger design increases comfort. Powder-Free to reduce risk of powder induced allergies, adhesion and granulomas.



**Surgical Glove
Sterile**



**DermAssist® Prestige®
DHD**

**Latex
Series 139**

Features:

- Unique Hydrogel Interior for Damp-hand Donnability
- Bisque Finish for an Improved Wet/Dry Grip
- Low Protein for Reduced Chances of Latex Allergic Reactions

PRODUCT DETAILS

SIZE	ITEM NO.	PACKAGING	DESCRIPTION
5.5	139550	50 pairs/box, 4 boxes/case	Gloves, Surgical, Latex, Sterile, Powder-Free, Bisque Finish, Damp Hand Don
6	139600	50 pairs/box, 4 boxes/case	
6.5	139650	50 pairs/box, 4 boxes/case	
7	139700	50 pairs/box, 4 boxes/case	
7.5	139750	50 pairs/box, 4 boxes/case	
8	139800	50 pairs/box, 4 boxes/case	
8.5	139850	50 pairs/box, 4 boxes/case	
9	139900	50 pairs/box, 4 boxes/case	

View this product on
our website:



Product Solutions You Trust




Specification Sheet

DermAssist®

Prestige®  **Smooth Donning**

One Sterile Pair **LATEX SURGICAL GLOVES**

Powder Free



Product Attributes

- Unique Hydrogel Interior
- Bisque Finish
- Low Protein

Benefits

- Damp-hand Donnability
- Improved Wet/Dry Grip
- Reduced Chances of Type I Allergic Reactions

DermAssist® Prestige® is manufactured in compliance with multiple international standards, including the following:

Designation	Standard
ASTM D3577	Standard Specification for Rubber Surgical Gloves
ASTM D5151	Standard Test Method for Detection of Holes in Medical Gloves
ASTM F1671	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens
ASTM D5712	Standard Test Method for Analysis of Aqueous Extractable Protein in Natural Rubber

Average Length	Average Palm Thickness	Average Finger Thickness
11.6 in ♦ 295 mm	7.5 mil ♦ 0.19 mm	8.7 mil ♦ 0.22 mm

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	29	24
ASTM Requirement Min. (Mpa)	24	18
Elongation (%)	850	820
ASTM Requirement Min. (%)	750	560



Intertek

Innovative Healthcare Corporation is certified to ISO 13485:2003 QMS for medical devices.



**INNOVATIVE
HEALTHCARE
CORPORATION**

Product Solutions You Trust

1-800-272-1533